

24 November 2016

## Advisory Notice – Faecal Pathogens A (16-well)

There will be a future update to a product you order. Please find the products, and details of the change below:

***Name of product affected: Faecal Pathogens A (16-well)***

***Previous catalogue number of product: 25031 Version 04***

***New catalogue number of product: 25031 Version 05***

***Regulatory status of updated product: IVD use in Australia and Europe (i.e. unchanged)***

***Reason for change:***

- The Salmonella has been updated to reduce the incidence of false positives. The gene target is unchanged.
- Noro 1 and Noro 2 assays updated to remove background. The gene target is unchanged.
- The RT conditions have been updated to those used in Respiratory and CSF panels. The gene target is unchanged.

***Date effective of change: 25 November 2016***

Please don't hesitate to contact us on 02 96988030 or at [support@ausdiagnostics.com](mailto:support@ausdiagnostics.com) if you have any further queries or require any assistance with the update procedure.