

20th June 2017

Advisory Notice: Melt range update – Noro-1

There is an update to a product you order. AusDiagnostics routine monitoring of melt temperature resulted in an update to the melt temperature range of the Noro-1 assay.

Please install the new template of the product you use from the AusDiagnostics website (www.ausdiagnostics.com) or ask your sales representative.

Please find the products affected below:

Name of IVD products affected:

- REF 25037 VER 02 - Enteric Viruses (8-well)
- REF 25031 VER 05 - Faecal Pathogens A (16-well)
- REF 25036 VER 05 - Faecal Pathogens C (16-well)
- REF 25039 VER 04 - Faecal Pathogens M (16-well)

Please don't hesitate to contact us on 02 96988030 or at support@ausdiagnostics.com if you have any further queries or require any assistance with the update procedure.



AusDiagnostics Pty Ltd
205 Victoria Street
BEACONSFIELD
NSW 2015 Australia
Phone: +612 9698 8030
Fax: +612 9319 1661



AusDiagnostics UK Ltd
Unit 3
Anglo Business Park, Asheridge Rd,
Chesham, Bucks, HP5 2QA UK
Tel: +44 (0) 1494 300121
Mob: +44 (0) 7896 534843