

18th May 2017

Advisory Notice: Melt change update – H3

There is an update to a product you order, AusDiagnostics routine monitoring of melt temperature resulted in an update to the melt temperature range of the flu typing assays. It should allow better calling of H3 and pdH1 positive samples.

Please install the new template for the product you use from the AusDiagnostics website (www.ausdiagnostics.com) or ask your sales representative.

Please find the products affected below:

Name of IVD products affected:

- REF 20602 VER 06 – Respiratory Viruses (16-well)
- REF 20602 VER 09 – Respiratory Viruses (16-well)
- REF 20616 VER 01 – Paediatric Respiratory Pathogens (16-well)
- REF 20613 VER 03 – Respiratory Pathogens C (16-well)

Please don't hesitate to contact us on 02 96988030 or at support@ausdiagnostics.com if you have any further queries or require any assistance with the update procedure.



AusDiagnostics Pty Ltd
205 Victoria Street
BEACONSFIELD
NSW 2015 Australia
Phone: +612 9698 8030
Fax: +612 9319 1661



AusDiagnostics UK Ltd
Unit 3
Anglo Business Park, Asheridge Rd,
Chesham, Bucks, HP5 2QA UK
Tel: +44 (0) 1494 300121
Mob: +44 (0) 7896 534843